
Midmark M11 - v1.0.0
Total Cycles: 3750

Name: RITTER 1
Sterilizer ID: V2178198
09/11/2023 12:40 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 170.6 F 0.0
2:00 172.6 F 2.6
4:00 196.0 F 7.1
6:00 214.0 F 11.5
8:00 224.1 F 6.1
10:00 234.5 F 11.1
12:00 233.8 F 8.4
14:00 242.0 F 11.2
16:00 246.0 F 13.2
18:00 257.9 F 19.5
20:00 268.8 F 26.3

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.2
1:00 272.3 F 28.7
2:00 271.4 F 28.5
3:00 272.1 F 28.6
4:00 271.2 F 28.3

Min 271.2 F 28.1
Max 272.6 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 190.4 F 0.0
5:00 216.2 F 0.0
10:00 228.6 F 0.0
15:00 230.2 F 0.0
20:00 227.8 F 0.0
25:00 228.9 F 0.0
30:00 227.4 F 0.0
Drying Complete

FILLING: 1:32
HEATING: 21:09
STERILIZING: 4:00
VENTING: 3:38
DRYING: 30:00
TOTAL CYCLE: 01:00:20
CYCLE COMPLETE

09/11/2023 01:40 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____