

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.1uS / 0.0ppm  
CYCLE NUMBER 002228  
11:21 06/11/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:13  
PRESSURIZING 2:19  
STERILIZING 3:28  
136.4 C 225kPa 3:28  
136.4 C 226kPa 3:58  
136.4 C 226kPa 4:28  
136.2 C 225kPa 4:58  
136.3 C 224kPa 5:28  
136.4 C 223kPa 5:58  
136.3 C 225kPa 6:28  
136.4 C 226kPa 6:58  
136.4 C 226kPa 6:58  
VENTING 6:58  
AIR DRYING 7:50  
CYCLE COMPLETE 23:17  
-----

Digital Signature #  
EC7A4FAD93C824A3  
-----