

-----  
Brampton, ON  
Canada  
-----

STATIM 5000 S501R713  
SN 511019D00004  
Unit # : 000  
WATER QUALITY  
1.2uS / 0.7ppm  
CYCLE NUMBER 001466  
13:54 09/30/2021  
UNWRAPPED  
134 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 4:27  
PRESSURIZING 5:58  
STERILIZING 7:33  
136.4 C 222kPa 7:33  
136.4 C 223kPa 8:03  
136.3 C 222kPa 8:33  
136.3 C 222kPa 9:03  
136.3 C 222kPa 9:33  
136.4 C 224kPa 10:03  
136.3 C 222kPa 10:33  
136.4 C 224kPa 11:03  
VENTING 11:03  
AIR DRYING 12:04  
CYCLE COMPLETE 28:56  
-----

Digital Signature #  
B0E0EEB75FFA9A2D  
-----