
Midmark M11 - v1.0.0
Total Cycles: 3198

Name: RITTER 1
Sterilizer ID: V2178198
05/03/2023 03:55 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 199.2 F 0.0
2:00 192.8 F 3.6
4:00 214.5 F 8.9
6:00 223.6 F 8.2
8:00 233.1 F 10.0
10:00 237.1 F 8.3
12:00 243.9 F 12.3
14:00 245.2 F 13.1
16:00 258.0 F 19.9
18:00 269.4 F 27.1

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 28.5
1:00 272.4 F 29.3
2:00 272.1 F 28.8
3:00 271.7 F 28.8
4:00 272.6 F 29.3

Min 271.5 F 28.4
Max 272.6 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.8 F 0.0
5:00 219.4 F 0.0
10:00 237.5 F 0.0
15:00 234.8 F 0.0
20:00 227.4 F 0.0
25:00 226.8 F 0.0
30:00 225.3 F 0.0
Drying Complete

FILLING: 2:03
HEATING: 18:54
STERILIZING: 4:00
VENTING: 3:25
DRYING: 30:00
TOTAL CYCLE: 00:58:23
CYCLE COMPLETE

05/03/2023 04:53 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____