
Midmark M11 - v1.0.0
Total Cycles: 4200

Name: RITTER 1
Sterilizer ID: V2178198
01/12/2024 03:09 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 93.7 F 0.0
2:00 98.4 F 0.8
4:00 111.2 F 2.8
6:00 141.6 F 5.6
8:00 166.0 F 8.6
10:00 185.4 F 12.4
12:00 208.9 F 4.5
14:00 219.8 F 8.2
16:00 229.9 F 12.2
18:00 230.5 F 7.9
20:00 241.7 F 12.8
22:00 239.6 F 10.4
24:00 250.9 F 16.0
26:00 261.4 F 21.9
28:00 271.1 F 28.3

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.5
1:00 272.0 F 29.1
2:00 272.1 F 28.9
3:00 271.1 F 28.5
4:00 272.4 F 29.3

Min 271.0 F 28.4
Max 272.7 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.2 F 0.0
5:00 218.2 F 0.0
10:00 219.9 F 0.0
15:00 226.6 F 0.0
20:00 232.4 F 0.0
25:00 226.9 F 0.0
30:00 224.1 F 0.0
Drying Complete

FILLING: 1:02
HEATING: 28:35
STERILIZING: 4:00
VENTING: 3:38
DRYING: 30:00
TOTAL CYCLE: 01:07:16
CYCLE COMPLETE

01/12/2024 04:16 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____