
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 004380
10:53 19/09/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:00
PRESSURIZING 2:02
STERILIZING 3:03
136.4 C 221kPa 3:03
136.4 C 221kPa 3:33
136.3 C 224kPa 4:03
136.2 C 223kPa 4:33
136.2 C 219kPa 5:03
136.4 C 223kPa 5:33
136.4 C 220kPa 6:03
136.1 C 219kPa 6:33
136.1 C 219kPa 6:33
VENTING 6:33
AIR DRYING 7:21
CYCLE COMPLETE 7:25

Digital Signature #
F56FB7FE0DAE8B87
