
Midmark M11 - v1.0.0
Total Cycles: 4297

Name: RITTER 1
Sterilizer ID: V2178198
02/13/2024 10:24 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 138.7 F 0.0
2:00 135.1 F 1.3
4:00 145.3 F 3.9
6:00 182.7 F 7.9
8:00 196.0 F 12.7
10:00 212.2 F 4.9
12:00 225.3 F 9.2
14:00 236.7 F 11.0
16:00 238.8 F 10.4
18:00 235.2 F 8.1
20:00 248.4 F 14.5
22:00 260.0 F 20.9
24:00 270.6 F 27.6

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.5
1:00 272.4 F 28.7
2:00 272.8 F 29.1
3:00 271.6 F 28.7
4:00 272.4 F 28.8

Min 271.5 F 28.4
Max 272.9 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 185.2 F 0.0
5:00 235.5 F 0.0
10:00 219.7 F 0.0
15:00 231.7 F 0.0
20:00 223.7 F 0.0
25:00 231.4 F 0.0
30:00 228.6 F 0.0
Drying Complete

FILLING: 0:58
HEATING: 25:29
STERILIZING: 4:00
VENTING: 3:25
DRYING: 30:00
TOTAL CYCLE: 01:03:52
CYCLE COMPLETE

02/13/2024 11:27 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____