

-----  
Midmark M11 - v1.0.5  
Total Cycles: 4257

Name: RITTER2  
Sterilizer ID: V2313228  
01/10/2025 02:05 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 119.1 F 0.0  
2:00 115.6 F 0.0  
4:00 136.6 F 0.2  
6:00 167.9 F 0.5  
8:00 192.1 F 3.0  
10:00 207.4 F 6.7  
12:00 220.9 F 10.9  
14:00 226.2 F 6.5  
16:00 238.1 F 12.0  
18:00 237.8 F 10.1  
20:00 237.8 F 9.1  
22:00 250.9 F 16.0  
24:00 262.6 F 22.8  
26:00 272.6 F 29.2

STERILIZING  
mm:ss Degrees PSI  
0:00 272.2 F 28.8  
1:00 272.4 F 29.3  
2:00 272.6 F 29.6  
3:00 271.7 F 29.0  
4:00 272.1 F 29.1

Min 271.6 F 28.7  
Max 272.7 F 29.6

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 183.9 F 0.0  
5:00 229.6 F 0.0  
10:00 209.5 F 0.0  
15:00 183.8 F 0.0  
20:00 179.3 F 0.0  
25:00 177.1 F 0.0  
30:00 175.6 F 0.0  
Drying Complete

FILLING: 1:01  
HEATING: 26:07  
STERILIZING: 4:00  
VENTING: 3:44  
DRYING: 30:00  
TOTAL CYCLE: 01:04:52  
CYCLE COMPLETE

01/10/2025 03:09 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_