

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3514

Name: RITTER 1  
Sterilizer ID: V2178198  
07/11/2023 01:26 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 198.8 F 0.0  
2:00 192.2 F 3.6  
4:00 212.5 F 8.8  
6:00 226.7 F 9.1  
8:00 234.4 F 10.4  
10:00 233.2 F 7.8  
12:00 241.0 F 10.6  
14:00 247.4 F 14.0  
16:00 260.2 F 21.1  
18:00 271.6 F 28.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 28.4  
1:00 272.5 F 28.9  
2:00 271.8 F 28.7  
3:00 272.3 F 28.8  
4:00 271.8 F 28.7  
  
Min 271.5 F 28.2  
Max 272.8 F 29.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 211.8 F 0.0  
5:00 220.6 F 0.0  
10:00 223.4 F 0.0  
15:00 227.1 F 0.0  
20:00 229.4 F 0.0  
25:00 229.3 F 0.0  
30:00 227.5 F 0.0  
Drying Complete

FILLING: 2:13  
HEATING: 18:35  
STERILIZING: 4:00  
VENTING: 3:13  
DRYING: 30:00  
TOTAL CYCLE: 00:58:02  
CYCLE COMPLETE

07/11/2023 02:23 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_