
Midmark M11 - v1.0.5
Total Cycles: 3915

Name: RITTER2
Sterilizer ID: V2313228
09/09/2024 09:33 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 68.6 F 0.0
2:00 72.0 F 0.0
4:00 95.3 F 0.0
6:00 97.2 F 0.0
8:00 145.9 F 0.1
10:00 176.8 F 0.4
12:00 197.8 F 1.2
14:00 201.6 F 4.2
16:00 215.9 F 7.2
18:00 227.5 F 10.6
20:00 231.8 F 8.7
22:00 236.8 F 10.7
24:00 235.5 F 8.6
26:00 243.5 F 12.0
28:00 245.5 F 13.2
30:00 256.6 F 19.2
32:00 266.7 F 25.4

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.7
1:00 272.4 F 29.1
2:00 271.7 F 29.1
3:00 272.1 F 29.0
4:00 271.8 F 29.2

Min 271.4 F 28.5
Max 272.6 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 198.2 F 0.0
5:00 212.4 F 0.0
10:00 202.1 F 0.0
15:00 193.9 F 0.0
20:00 200.0 F 0.0
25:00 200.5 F 0.0
30:00 204.9 F 0.0

Drying Complete

FILLING: 0:51
HEATING: 33:37
STERILIZING: 4:00
VENTING: 3:33
DRYING: 30:00
TOTAL CYCLE: 01:12:00
CYCLE COMPLETE

09/09/2024 10:44 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____