
Surrey, BC
Canada

STATIM 5000 S501R711
SN 511018F00076

Unit # : 000

WATER QUALITY

3.3uS / 2.1ppm

CYCLE NUMBER 000009

10:15 13/01/2020

RUBBER AND PLASTICS

121 C for 15 min.

CYCLE START 0:00

CONDITIONING 3:32

PRESSURIZING 9:38

STERILIZING 10:38

123.4 C 120kPa 10:38

123.4 C 120kPa 11:08

123.5 C 120kPa 11:38

123.4 C 120kPa 12:08

123.4 C 120kPa 12:38

123.5 C 120kPa 13:08

123.4 C 120kPa 13:38

123.4 C 120kPa 14:08

123.2 C 120kPa 14:38

123.4 C 119kPa 15:08

123.4 C 120kPa 15:38

123.4 C 118kPa 16:08

123.4 C 120kPa 16:38

123.5 C 120kPa 17:09

123.4 C 120kPa 17:39

123.4 C 120kPa 18:09

123.5 C 120kPa 18:39

123.4 C 120kPa 19:09

123.4 C 120kPa 19:39

123.5 C 120kPa 20:09

123.4 C 120kPa 20:39

123.3 C 120kPa 21:09

123.3 C 119kPa 21:39

123.5 C 120kPa 22:09

123.4 C 120kPa 22:40

123.4 C 120kPa 23:10

123.3 C 120kPa 23:40

123.4 C 120kPa 24:10

123.5 C 120kPa 24:40

123.4 C 121kPa 25:10

123.4 C 120kPa 25:38

VENTING 25:38

AIR DRYING 26:43
CYCLE COMPLETE 56:43

Digital Signature #
90273FFD0F6E78CF
