
Midmark M11 - v1.0.5
Total Cycles: 3207

Name: RITTER2
Sterilizer ID: V2313228
12/12/2023 10:41 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 70.0 F 0.0
2:00 70.2 F 0.0
4:00 72.3 F 0.2
6:00 96.0 F 1.0
8:00 119.5 F 3.1
10:00 145.5 F 5.9
12:00 186.7 F 8.7
14:00 202.5 F 11.9
16:00 209.6 F 4.5
18:00 217.6 F 8.2
20:00 229.1 F 12.0
22:00 229.4 F 7.9
24:00 241.0 F 12.8
26:00 239.2 F 10.8
28:00 250.7 F 16.4
30:00 261.0 F 22.3
32:00 270.5 F 28.7

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 29.7
1:00 272.7 F 30.0
2:00 272.0 F 30.1
3:00 272.4 F 29.8
4:00 272.4 F 30.3

Min 271.6 F 29.5
Max 272.9 F 30.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.5 F 0.0
5:00 220.6 F 0.0
10:00 221.5 F 0.0
15:00 216.6 F 0.0
20:00 230.4 F 0.0
25:00 229.3 F 0.0
30:00 226.9 F 0.0

Drying Complete

FILLING: 0:50
HEATING: 33:35
STERILIZING: 4:00
VENTING: 3:33
DRYING: 30:00
TOTAL CYCLE: 01:11:58
CYCLE COMPLETE

12/12/2023 11:52 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____