
Midmark M11 - v1.0.0
Total Cycles: 1799

Name: RITTER 1
Sterilizer ID: V2178198
04/06/2022 11:14 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 174.0 F 0.0
2:00 165.4 F 2.7
4:00 175.2 F 6.7
6:00 197.7 F 11.7
8:00 217.7 F 5.2
10:00 233.4 F 10.1
12:00 235.4 F 8.2
14:00 243.4 F 12.3
16:00 240.8 F 10.6
18:00 253.7 F 17.1
20:00 264.9 F 23.8

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.2
1:00 272.4 F 28.7
2:00 271.5 F 28.4
3:00 272.2 F 28.7
4:00 271.4 F 28.4

Min 271.3 F 28.1
Max 272.7 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 192.8 F 0.0
5:00 228.6 F 0.0
10:00 192.6 F 0.0
15:00 187.9 F 0.0
20:00 186.3 F 0.0
25:00 189.4 F 0.0
30:00 202.3 F 0.0
Drying Complete

FILLING: 1:05
HEATING: 21:54
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 01:00:11
CYCLE COMPLETE

04/06/2022 12:14 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____