

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1203

Name: RITTER 1  
Sterilizer ID: V2178198  
09/27/2021 06:40 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 70.7 F 0.0  
2:00 73.2 F 0.0  
4:00 99.0 F 0.0  
6:00 116.0 F 0.1  
8:00 149.9 F 0.2  
10:00 175.1 F 0.2  
12:00 197.5 F 1.9  
14:00 209.1 F 5.3  
16:00 218.7 F 8.6  
18:00 230.1 F 12.2  
20:00 230.4 F 7.7  
22:00 242.3 F 12.9  
24:00 240.1 F 10.7  
26:00 237.1 F 9.1  
28:00 251.5 F 16.0  
30:00 263.0 F 22.7  
32:00 272.2 F 28.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.7 F 28.3  
1:00 272.7 F 28.8  
2:00 271.9 F 28.5  
3:00 272.6 F 28.7  
4:00 271.9 F 28.5  
  
Min 271.7 F 28.1  
Max 273.0 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 180.3 F 0.0  
5:00 196.4 F 0.0  
10:00 203.7 F 0.0  
15:00 199.7 F 0.0  
20:00 186.2 F 0.0  
25:00 190.9 F 0.0  
30:00 188.9 F 0.0  
  
Drying Complete

FILLING: 0:50  
HEATING: 32:11  
STERILIZING: 4:00  
VENTING: 3:03  
DRYING: 30:00  
TOTAL CYCLE: 01:10:05  
CYCLE COMPLETE

09/27/2021 07:50 AM  
ACCEPTED: X  
REJECTED:  
Operator: U01