

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.1uS / 0.0ppm  
CYCLE NUMBER 005337  
11:51 10/04/2023  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 1:53  
PRESSURIZING 1:58  
STERILIZING 2:51  
136.4 C 220kPa 2:51  
136.4 C 224kPa 3:21  
136.1 C 219kPa 3:51  
136.3 C 219kPa 4:21  
136.4 C 223kPa 4:51  
136.4 C 221kPa 5:21  
136.3 C 222kPa 5:51  
136.2 C 224kPa 6:21  
136.2 C 224kPa 6:21  
VENTING 6:21  
AIR DRYING 7:13  
CYCLE COMPLETE 12:14

-----  
Digital Signature #  
BC55EAD87260EB61  
-----