
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 001859
10:03 16/07/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:12
PRESSURIZING 3:16
STERILIZING 4:21
136.0 C 221kPa 4:21
136.3 C 225kPa 4:51
136.2 C 224kPa 5:21
136.3 C 224kPa 5:51
136.1 C 222kPa 6:21
136.4 C 226kPa 6:51
136.4 C 225kPa 7:21
136.4 C 224kPa 7:51
136.4 C 224kPa 7:51
VENTING 7:51
AIR DRYING 8:47
CYCLE COMPLETE 8:49

Digital Signature #
63C78CDA050E47DC
