

-----

Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.4uS / 0.2ppm  
CYCLE NUMBER 002902  
13:28 29/12/2021  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 1:55  
PRESSURIZING 1:57  
STERILIZING 2:52  
136.5 C 220kPa 2:52  
136.1 C 224kPa 3:22  
136.4 C 222kPa 3:52  
136.2 C 221kPa 4:22  
136.4 C 222kPa 4:52  
136.4 C 225kPa 5:22  
136.1 C 222kPa 5:52  
136.4 C 224kPa 6:22  
136.4 C 224kPa 6:22  
VENTING 6:22  
AIR DRYING 7:58  
CYCLE COMPLETE 8:13  
-----

Digital Signature #  
5CDC6D6559AC43C0  
-----