
Midmark M11 - v1.0.5
Total Cycles: 593

Name: RITTER2
Sterilizer ID: V2313228
12/01/2021 09:40 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 111.5 F 0.0
2:00 110.7 F 0.8
4:00 121.6 F 3.0
6:00 151.9 F 6.5
8:00 180.4 F 11.1
10:00 213.3 F 3.5
12:00 220.0 F 8.2
14:00 231.8 F 12.9
16:00 234.0 F 9.7
18:00 237.8 F 9.3
20:00 246.7 F 14.3
22:00 258.5 F 20.8
24:00 269.6 F 27.8

STERILIZING
mm:ss Degrees PSI
0:00 272.4 F 29.1
1:00 272.0 F 29.3
2:00 272.7 F 29.5
3:00 272.2 F 29.0
4:00 271.9 F 29.3

Min 271.8 F 28.9
Max 272.9 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 181.2 F 0.0
5:00 241.1 F 0.0
10:00 224.7 F 0.0
15:00 186.3 F 0.0
20:00 181.9 F 0.0
25:00 179.0 F 0.0
30:00 177.2 F 0.0
Drying Complete

FILLING: 1:07
HEATING: 24:40
STERILIZING: 4:00
VENTING: 3:01
DRYING: 30:00
TOTAL CYCLE: 01:02:48
CYCLE COMPLETE

12/01/2021 10:45 AM
ACCEPTED: X
REJECTED:
Operator: U02