
Midmark M11 - v1.0.0
Total Cycles: 4442

Name: RITTER 1
Sterilizer ID: V2178198
04/10/2024 09:51 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 160.4 F 0.0
2:00 155.1 F 1.7
4:00 155.1 F 4.5
6:00 185.2 F 8.0
8:00 198.8 F 12.5
10:00 213.9 F 5.3
12:00 222.5 F 9.6
14:00 231.8 F 10.0
16:00 237.5 F 10.8
18:00 234.7 F 8.3
20:00 248.3 F 14.7
22:00 259.2 F 20.7
24:00 269.3 F 27.2

STERILIZING
mm:ss Degrees PSI
0:00 272.0 F 29.0
1:00 272.7 F 29.6
2:00 271.8 F 28.9
3:00 272.8 F 29.6
4:00 271.9 F 28.9

Min 271.6 F 28.8
Max 272.9 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.7 F 0.0
5:00 221.9 F 0.0
10:00 230.2 F 0.0
15:00 226.6 F 0.0
20:00 223.6 F 0.0
25:00 221.8 F 0.0
30:00 220.6 F 0.0
Drying Complete

FILLING: 1:00
HEATING: 25:47
STERILIZING: 4:00
VENTING: 3:41
DRYING: 30:00
TOTAL CYCLE: 01:04:29
CYCLE COMPLETE

04/10/2024 10:55 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____