

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 002996  
10:44 20/01/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:33  
PRESSURIZING 2:36  
STERILIZING 3:26  
136.0 C 220kPa 3:26  
136.4 C 224kPa 3:56  
136.4 C 225kPa 4:26  
136.4 C 224kPa 4:56  
136.0 C 221kPa 5:26  
136.4 C 224kPa 5:56  
136.4 C 221kPa 6:26  
136.2 C 222kPa 6:56  
136.2 C 222kPa 6:56  
VENTING 6:56  
AIR DRYING 7:50  
CYCLE COMPLETE 8:12

-----  
Digital Signature #  
28006343C7CB8F68  
-----