
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 005451
12:47 01/05/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:46
PRESSURIZING 1:51
STERILIZING 2:35
136.0 C 220kPa 2:35
136.3 C 221kPa 3:05
136.4 C 220kPa 3:35
136.4 C 222kPa 4:05
136.4 C 224kPa 4:35
136.4 C 224kPa 5:05
136.1 C 220kPa 5:35
136.4 C 221kPa 6:05
136.4 C 221kPa 6:05
VENTING 6:05
AIR DRYING 6:55
CYCLE COMPLETE 16:57

Digital Signature #
D182908D3C8C8147
