
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 007408
7:42 22/05/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:24
PRESSURIZING 2:28
STERILIZING 3:35
136.0 C 220kPa 3:35
136.2 C 219kPa 4:05
136.3 C 221kPa 4:35
136.3 C 223kPa 5:05
136.3 C 222kPa 5:35
136.3 C 223kPa 6:05
136.1 C 221kPa 6:35
136.5 C 223kPa 7:05
136.5 C 223kPa 7:05
VENTING 7:05
AIR DRYING 7:53
CYCLE COMPLETE 37:53

Digital Signature #
18E771C6A232A097
