

-----  
Midmark M11 - v1.0.5  
Total Cycles: 555

Name: RITTER2  
Sterilizer ID: V2313228  
11/17/2021 07:32 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 72.2 F 0.0  
2:00 78.5 F 0.0  
4:00 118.9 F 0.0  
6:00 119.4 F 0.1  
8:00 157.0 F 0.3  
10:00 183.5 F 0.9  
12:00 199.4 F 4.7  
14:00 215.4 F 8.4  
16:00 226.3 F 12.3  
18:00 228.7 F 7.3  
20:00 239.5 F 12.1  
22:00 237.4 F 9.7  
24:00 238.5 F 9.3  
26:00 249.0 F 14.8  
28:00 260.5 F 21.3  
30:00 270.9 F 28.2

STERILIZING  
mm:ss Degrees PSI  
0:00 272.1 F 28.6  
1:00 272.8 F 29.1  
2:00 272.5 F 29.2  
3:00 271.7 F 28.5  
4:00 272.5 F 29.0  
  
Min 271.7 F 28.4  
Max 272.9 F 29.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 172.4 F 0.0  
5:00 210.6 F 0.0  
10:00 220.6 F 0.0  
15:00 213.5 F 0.0  
20:00 204.5 F 0.0  
25:00 201.4 F 0.0  
30:00 198.2 F 0.0  
Drying Complete

FILLING: 0:56  
HEATING: 30:32  
STERILIZING: 4:00  
VENTING: 2:57  
DRYING: 30:00  
TOTAL CYCLE: 01:08:25  
CYCLE COMPLETE

11/17/2021 08:40 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_