
Midmark M11 - v1.0.0
Total Cycles: 1891

Name: RITTER 1
Sterilizer ID: V2178198
05/02/2022 09:33 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 211.2 F 0.0
2:00 207.9 F 4.1
4:00 208.4 F 9.4
6:00 223.7 F 7.0
8:00 235.2 F 10.5
10:00 233.6 F 7.3
12:00 243.1 F 11.3
14:00 245.8 F 13.2
16:00 258.7 F 20.2
18:00 270.1 F 27.5

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.4
1:00 272.6 F 29.2
2:00 271.9 F 28.6
3:00 271.3 F 28.6
4:00 272.6 F 29.3

Min 271.2 F 28.3
Max 272.7 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.6 F 0.0
5:00 207.9 F 0.0
10:00 204.0 F 0.0
15:00 233.3 F 0.0
20:00 237.5 F 0.0
25:00 231.9 F 0.0
30:00 227.6 F 0.0

Drying Complete

FILLING: 2:56
HEATING: 18:56
STERILIZING: 4:00
VENTING: 3:07
DRYING: 30:00
TOTAL CYCLE: 00:58:59
CYCLE COMPLETE

05/02/2022 10:31 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____