

-----  
Midmark M11 - v1.0.5  
Total Cycles: 478

Name: RITTER2  
Sterilizer ID: V2313228  
10/25/2021 08:36 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 112.3 F 0.0  
2:00 114.1 F 0.8  
4:00 123.3 F 2.7  
6:00 144.3 F 6.0  
8:00 178.2 F 10.9  
10:00 211.3 F 5.0  
12:00 222.8 F 10.5  
14:00 227.9 F 7.3  
16:00 239.3 F 12.8  
18:00 239.6 F 11.1  
20:00 252.2 F 17.5  
22:00 263.5 F 24.2  
24:00 271.4 F 29.3

STERILIZING  
mm:ss Degrees PSI  
0:00 272.5 F 30.1  
1:00 272.5 F 30.1  
2:00 272.6 F 30.2  
3:00 272.8 F 30.3  
4:00 272.7 F 30.5

Min 272.0 F 30.0  
Max 273.0 F 30.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 204.5 F 0.0  
5:00 210.5 F 0.0  
10:00 251.2 F 0.0  
15:00 237.6 F 0.0  
20:00 224.0 F 0.0  
25:00 224.6 F 0.0  
30:00 187.7 F 0.0  
Drying Complete

FILLING: 0:53  
HEATING: 24:40  
STERILIZING: 4:00  
VENTING: 2:57  
DRYING: 30:00  
TOTAL CYCLE: 01:02:30  
CYCLE COMPLETE

10/25/2021 10:00 AM  
ACCEPTED: X  
REJECTED:  
Operator: U01