

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4357

Name: RITTER 1  
Sterilizer ID: V2178198  
03/04/2024 02:46 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 90.3 F 0.0  
2:00 90.8 F 0.6  
4:00 99.7 F 2.2  
6:00 126.5 F 4.7  
8:00 158.1 F 7.5  
10:00 178.1 F 10.8  
12:00 208.1 F 4.6  
14:00 210.9 F 5.5  
16:00 220.6 F 9.0  
18:00 230.6 F 12.8  
20:00 232.2 F 8.4  
22:00 243.2 F 13.3  
24:00 241.7 F 11.3  
26:00 253.0 F 17.0  
28:00 263.1 F 22.9  
30:00 272.7 F 29.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 28.4  
1:00 272.3 F 29.2  
2:00 272.0 F 28.7  
3:00 271.1 F 28.4  
4:00 272.2 F 29.2

Min 271.1 F 28.2  
Max 272.6 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 185.0 F 0.0  
5:00 221.9 F 0.0  
10:00 229.5 F 0.0  
15:00 229.3 F 0.0  
20:00 229.8 F 0.0  
25:00 230.1 F 0.0  
30:00 227.3 F 0.0  
Drying Complete

FILLING: 0:55  
HEATING: 30:27  
STERILIZING: 4:00  
VENTING: 3:45  
DRYING: 30:00  
TOTAL CYCLE: 01:09:07  
CYCLE COMPLETE

03/04/2024 03:55 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_