
Midmark M11 - v1.0.5
Total Cycles: 1779

Name: RITTER2
Sterilizer ID: V2313228
11/28/2022 02:39 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 184.2 F 0.0
2:00 182.0 F 2.3
4:00 190.8 F 7.0
6:00 209.8 F 12.6
8:00 225.5 F 8.5
10:00 236.2 F 10.7
12:00 241.4 F 12.7
14:00 242.2 F 12.1
16:00 255.8 F 19.2
18:00 267.6 F 26.6

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 29.3
1:00 272.6 F 29.8
2:00 271.8 F 29.6
3:00 272.5 F 29.8
4:00 271.6 F 29.4

Min 271.6 F 29.2
Max 272.8 F 30.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.7 F 0.0
5:00 223.9 F 0.0
10:00 237.0 F 0.0
15:00 231.8 F 0.0
20:00 224.5 F 0.0
25:00 222.1 F 0.0
30:00 221.5 F 0.0
Drying Complete

FILLING: 2:01
HEATING: 19:18
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 00:58:31
CYCLE COMPLETE

11/28/2022 03:37 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____