
Midmark M11 - v1.0.0
Total Cycles: 3990

Name: RITTER 1
Sterilizer ID: V2178198
11/07/2023 01:17 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 122.8 F 0.0
2:00 125.8 F 1.5
4:00 147.9 F 4.5
6:00 174.1 F 8.4
8:00 195.7 F 12.7
10:00 217.1 F 5.5
12:00 228.7 F 10.0
14:00 232.8 F 8.2
16:00 239.8 F 11.1
18:00 235.6 F 8.6
20:00 249.1 F 14.9
22:00 260.7 F 21.2
24:00 271.2 F 28.1

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 28.3
1:00 272.4 F 29.1
2:00 271.6 F 28.3
3:00 272.4 F 29.1
4:00 272.0 F 28.6

Min 271.1 F 28.2
Max 272.7 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 187.7 F 0.0
5:00 223.7 F 0.0
10:00 229.6 F 0.0
15:00 225.3 F 0.0
20:00 227.4 F 0.0
25:00 230.5 F 0.0
30:00 228.2 F 0.0
Drying Complete

FILLING: 1:01
HEATING: 24:33
STERILIZING: 4:00
VENTING: 3:24
DRYING: 30:00
TOTAL CYCLE: 01:02:59
CYCLE COMPLETE

11/07/2023 02:19 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____