

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2299

Name: RITTER2  
Sterilizer ID: V2313228  
04/25/2023 11:02 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 196.2 F 0.0  
2:00 188.4 F 4.1  
4:00 201.5 F 9.5  
6:00 219.7 F 6.8  
8:00 232.6 F 11.1  
10:00 234.4 F 9.1  
12:00 237.8 F 9.5  
14:00 249.9 F 15.6  
16:00 262.5 F 22.9  
18:00 272.4 F 29.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 29.0  
1:00 272.8 F 29.7  
2:00 271.9 F 29.0  
3:00 271.8 F 29.5  
4:00 272.6 F 29.6  
  
Min 271.3 F 28.8  
Max 272.8 F 29.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 204.9 F 0.0  
5:00 218.4 F 0.0  
10:00 226.9 F 0.0  
15:00 232.8 F 0.0  
20:00 225.7 F 0.0  
25:00 224.9 F 0.0  
30:00 219.3 F 0.0  
Drying Complete

FILLING: 1:55  
HEATING: 18:14  
STERILIZING: 4:00  
VENTING: 3:14  
DRYING: 30:00  
TOTAL CYCLE: 00:57:22  
CYCLE COMPLETE

04/25/2023 11:59 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_