

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.5uS / 0.3ppm  
CYCLE NUMBER 003050  
18:22 31/01/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:58  
PRESSURIZING 2:59  
STERILIZING 4:08  
136.5 C 225kPa 4:08  
136.3 C 224kPa 4:38  
136.3 C 223kPa 5:08  
136.2 C 224kPa 5:38  
136.3 C 223kPa 6:08  
136.4 C 223kPa 6:38  
136.4 C 225kPa 7:08  
136.4 C 224kPa 7:38  
136.4 C 224kPa 7:38  
VENTING 7:38  
AIR DRYING 8:35  
CYCLE COMPLETE 8:37

-----  
Digital Signature #  
2501A1344FF5A1A8  
-----