
Midmark M11 - v1.0.0
Total Cycles: 1291

Name: RITTER 1
Sterilizer ID: V2178198
10/22/2021 12:07 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 169.6 F 0.0
2:00 168.9 F 2.5
4:00 190.8 F 6.0
6:00 210.8 F 11.5
8:00 220.7 F 5.5
10:00 232.4 F 10.4
12:00 233.9 F 8.2
14:00 243.6 F 13.0
16:00 242.2 F 11.6
18:00 255.1 F 18.1
20:00 266.2 F 25.0

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.5
1:00 272.4 F 29.0
2:00 271.6 F 28.5
3:00 272.5 F 29.2
4:00 271.7 F 28.4

Min 271.5 F 28.4
Max 272.7 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.8 F 0.0
5:00 218.1 F 0.0
10:00 220.4 F 0.0
15:00 225.3 F 0.0
20:00 226.1 F 0.0
25:00 226.8 F 0.0
30:00 223.3 F 0.0
Drying Complete

FILLING: 1:37
HEATING: 21:37
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 01:00:28
CYCLE COMPLETE

10/22/2021 01:08 PM
ACCEPTED: X
REJECTED:
Operator: U02