

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1943

Name: RITTER2  
Sterilizer ID: V2313228  
01/18/2023 09:35 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 185.6 F 0.0  
2:00 183.5 F 2.5  
4:00 182.6 F 6.6  
6:00 203.5 F 12.1  
8:00 220.3 F 6.4  
10:00 234.7 F 12.0  
12:00 236.5 F 9.9  
14:00 236.8 F 8.9  
16:00 250.3 F 15.8  
18:00 262.6 F 22.9  
20:00 272.2 F 29.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.9 F 28.8  
1:00 272.4 F 29.6  
2:00 272.1 F 28.9  
3:00 272.0 F 29.3  
4:00 272.4 F 29.3

Min 271.4 F 28.6  
Max 272.8 F 29.6

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 195.5 F 0.0  
5:00 224.0 F 0.0  
10:00 220.9 F 0.0  
15:00 211.8 F 0.0  
20:00 207.6 F 0.0  
25:00 210.0 F 0.0  
30:00 223.4 F 0.0  
Drying Complete

FILLING: 1:28  
HEATING: 20:04  
STERILIZING: 4:00  
VENTING: 3:05  
DRYING: 30:00  
TOTAL CYCLE: 00:58:37  
CYCLE COMPLETE

01/18/2023 10:33 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_