
Midmark M11 - v1.0.0
Total Cycles: 685

Name: MIDMARK
Sterilizer ID: V2178198
02/22/2021 10:05 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 169.6 F 0.0
2:00 165.4 F 1.9
4:00 164.8 F 5.4
6:00 186.4 F 10.1
8:00 218.4 F 5.4
10:00 222.8 F 10.5
12:00 229.3 F 7.0
14:00 245.9 F 13.2
16:00 242.4 F 11.4
18:00 255.0 F 17.8
20:00 266.1 F 24.5
22:00 272.9 F 29.1

STERILIZING
mm:ss Degrees PSI
0:00 272.0 F 28.7
1:00 272.3 F 29.1
2:00 272.6 F 28.8
3:00 272.2 F 28.6
4:00 271.8 F 28.8

Min 271.7 F 28.6
Max 272.7 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 200.5 F 0.0
5:00 237.6 F 0.0
10:00 228.5 F 0.0
15:00 228.5 F 0.0
20:00 229.2 F 0.0
25:00 228.5 F 0.0
30:00 203.1 F 0.0
Drying Complete

FILLING: 1:08
HEATING: 22:16
STERILIZING: 4:00
VENTING: 3:02
DRYING: 30:00
TOTAL CYCLE: 01:00:26
CYCLE COMPLETE

02/22/2021 11:07 AM
ACCEPTED: X
REJECTED:
Operator: U01