

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 002026  
8:25 14/07/2021  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 1:56  
PRESSURIZING 2:01  
STERILIZING 3:00  
136.5 C 224kPa 3:00  
136.4 C 224kPa 3:30  
136.1 C 221kPa 4:00  
136.4 C 224kPa 4:30  
136.3 C 222kPa 5:00  
136.4 C 221kPa 5:30  
136.1 C 221kPa 6:00  
136.4 C 224kPa 6:30  
136.4 C 224kPa 6:30  
VENTING 6:31  
AIR DRYING 7:22  
CYCLE COMPLETE 7:25

-----  
Digital Signature #  
8E4A57992F4803E0  
-----