
Midmark M11 - v1.0.5
Total Cycles: 2023

Name: RITTER2
Sterilizer ID: V2313228
02/08/2023 09:51 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 170.5 F 0.0
2:00 174.9 F 0.1
4:00 173.1 F 1.4
6:00 192.9 F 5.8
8:00 222.7 F 10.8
10:00 226.5 F 6.9
12:00 238.2 F 12.5
14:00 238.1 F 10.6
16:00 237.0 F 9.5
18:00 251.2 F 16.5
20:00 263.4 F 23.5
22:00 272.0 F 29.0

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.0
1:00 272.6 F 29.8
2:00 272.0 F 29.1
3:00 271.6 F 29.2
4:00 272.6 F 29.8

Min 271.4 F 28.8
Max 272.7 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.8 F 0.0
5:00 223.3 F 0.0
10:00 238.1 F 0.0
15:00 223.4 F 0.0
20:00 220.4 F 0.0
25:00 220.6 F 0.0
30:00 219.7 F 0.0
Drying Complete

FILLING: 0:53
HEATING: 22:10
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 01:00:12
CYCLE COMPLETE

02/08/2023 10:51 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____