
Midmark M11 - v1.0.5
Total Cycles: 2387

Name: RITTER2
Sterilizer ID: V2313228
05/18/2023 09:06 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 140.4 F 0.0
2:00 139.8 F 0.1
4:00 154.4 F 0.2
6:00 183.3 F 3.9
8:00 201.4 F 8.3
10:00 221.9 F 13.2
12:00 226.1 F 8.9
14:00 236.1 F 10.5
16:00 241.2 F 12.3
18:00 241.0 F 11.2
20:00 254.4 F 18.1
22:00 265.9 F 25.2

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 29.0
1:00 271.3 F 29.3
2:00 272.4 F 29.8
3:00 273.1 F 30.1
4:00 272.1 F 29.5

Min 271.0 F 28.9
Max 273.6 F 30.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 187.3 F 0.0
5:00 209.4 F 0.0
10:00 225.1 F 0.0
15:00 231.0 F 0.0
20:00 231.7 F 0.0
25:00 231.2 F 0.0
30:00 227.4 F 0.0
Drying Complete

FILLING: 0:54
HEATING: 23:35
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 01:01:40
CYCLE COMPLETE

05/18/2023 10:07 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____