

-----  
Mississauga, ON  
Canada  
-----

STATIM 5000 S501R751  
SN 511020J00004  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 000795  
14:11 16/03/2022  
UNWRAPPED  
134 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 4:32  
PRESSURIZING 6:23  
STERILIZING 7:52  
136.0 C 220kPa 7:52  
136.1 C 218kPa 8:22  
135.8 C 218kPa 8:52  
136.3 C 221kPa 9:22  
136.1 C 221kPa 9:52  
136.3 C 222kPa 10:22  
136.4 C 221kPa 10:52  
136.4 C 223kPa 11:22  
136.4 C 223kPa 11:22  
VENTING 11:22  
AIR DRYING 12:25  
CYCLE COMPLETE 28:05  
-----

Digital Signature #  
E185B07346BE01C2  
-----