
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 006511
13:54 20/10/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:18
PRESSURIZING 2:22
STERILIZING 3:19
136.4 C 223kPa 3:19
136.4 C 222kPa 3:49
136.4 C 224kPa 4:19
136.4 C 222kPa 4:49
136.5 C 221kPa 5:19
136.4 C 224kPa 5:49
136.4 C 224kPa 6:19
136.4 C 224kPa 6:49
136.4 C 224kPa 6:49
VENTING 6:49
AIR DRYING 7:41
CYCLE COMPLETE 37:41

Digital Signature #
150C69D7F4001C3F
