
Midmark M11 - v1.0.5
Total Cycles: 3553

Name: RITTER2
Sterilizer ID: V2313228
04/24/2024 11:37 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 128.7 F 0.0
2:00 130.8 F 1.5
4:00 144.7 F 3.9
6:00 165.8 F 7.1
8:00 186.6 F 10.8
10:00 213.7 F 5.6
12:00 223.0 F 8.2
14:00 237.8 F 12.6
16:00 233.5 F 9.2
18:00 239.6 F 10.7
20:00 244.6 F 13.0
22:00 255.5 F 18.9
24:00 265.2 F 25.2
26:00 271.7 F 29.6

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 29.5
1:00 272.5 F 30.1
2:00 272.4 F 29.8
3:00 271.4 F 29.6
4:00 272.5 F 30.2

Min 271.3 F 29.4
Max 273.2 F 30.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 177.6 F 0.0
5:00 238.1 F 0.0
10:00 212.2 F 0.0
15:00 183.1 F 0.0
20:00 178.8 F 0.0
25:00 177.9 F 0.0
30:00 177.5 F 0.0

Drying Complete

FILLING: 0:58
HEATING: 26:31
STERILIZING: 4:00
VENTING: 3:37
DRYING: 30:00
TOTAL CYCLE: 01:05:06
CYCLE COMPLETE

04/24/2024 12:42 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____