
Midmark M11 - v1.0.5
Total Cycles: 2638

Name: RITTER2
Sterilizer ID: V2313228
07/12/2023 12:59 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 191.7 F 0.0
2:00 186.6 F 3.2
4:00 192.5 F 7.8
6:00 209.6 F 13.0
8:00 225.5 F 8.1
10:00 237.0 F 13.4
12:00 238.4 F 11.1
14:00 237.4 F 9.8
16:00 251.6 F 16.5
18:00 263.2 F 23.4

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.9
1:00 271.7 F 29.6
2:00 273.0 F 30.1
3:00 272.0 F 29.5
4:00 272.1 F 30.0

Min 270.9 F 28.8
Max 273.1 F 30.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 201.0 F 0.0
5:00 222.5 F 0.0
10:00 225.8 F 0.0
15:00 222.1 F 0.0
20:00 220.5 F 0.0
25:00 220.7 F 0.0
30:00 220.2 F 0.0
Drying Complete

FILLING: 1:49
HEATING: 19:59
STERILIZING: 4:00
VENTING: 3:16
DRYING: 30:00
TOTAL CYCLE: 00:59:05
CYCLE COMPLETE

07/12/2023 01:58 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____