
Midmark M11 - v1.0.5
Total Cycles: 765

Name: RITTER2
Sterilizer ID: V2313228
02/07/2022 02:36 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 156.1 F 0.0
2:00 163.6 F 3.3
4:00 188.1 F 8.1
6:00 209.2 F 11.5
8:00 225.3 F 8.0
10:00 237.9 F 12.3
12:00 242.0 F 12.3
14:00 243.0 F 12.0
16:00 256.6 F 19.2
18:00 268.7 F 26.9

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.8
1:00 273.0 F 29.5
2:00 271.6 F 28.7
3:00 272.1 F 29.6
4:00 272.4 F 29.4

Min 271.4 F 28.7
Max 273.0 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 179.9 F 0.0
5:00 219.8 F 0.0
10:00 210.8 F 0.0
15:00 205.7 F 0.0
20:00 202.9 F 0.0
25:00 200.8 F 0.0
30:00 199.5 F 0.0
Drying Complete

FILLING: 1:32
HEATING: 19:09
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 00:57:52
CYCLE COMPLETE

02/07/2022 03:33 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____