

-----  
Midmark M11 - v1.0.5  
Total Cycles: 4430

Name: RITTER2  
Sterilizer ID: V2313228  
03/14/2025 03:08 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 79.2 F 0.0  
2:00 79.8 F 0.3  
4:00 87.1 F 1.6  
6:00 109.9 F 3.8  
8:00 137.1 F 6.7  
10:00 157.8 F 9.8  
12:00 180.2 F 13.1  
14:00 202.5 F 4.5  
16:00 216.5 F 7.7  
18:00 229.5 F 11.4  
20:00 229.6 F 6.9  
22:00 237.2 F 12.2  
24:00 237.6 F 10.0  
26:00 249.0 F 15.5  
28:00 258.2 F 21.1  
30:00 267.6 F 27.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.1 F 30.0  
1:00 271.2 F 30.2  
2:00 271.4 F 30.4  
3:00 271.6 F 30.6  
4:00 271.7 F 30.7

Min 271.0 F 29.9  
Max 273.7 F 31.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 196.0 F 0.0  
5:00 181.1 F 0.0  
10:00 195.3 F 0.0  
15:00 194.0 F 0.0  
20:00 194.0 F 0.0  
25:00 197.7 F 0.0  
30:00 202.9 F 0.0  
Drying Complete

FILLING: 0:54  
HEATING: 31:44  
STERILIZING: 4:00  
VENTING: 3:56  
DRYING: 30:00  
TOTAL CYCLE: 01:10:34  
CYCLE COMPLETE

03/14/2025 04:18 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_