

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1871

Name: RITTER2  
Sterilizer ID: V2313228  
12/21/2022 11:19 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 170.5 F 0.0  
2:00 170.5 F 2.7  
4:00 177.7 F 6.9  
6:00 198.4 F 12.2  
8:00 220.6 F 7.0  
10:00 235.3 F 12.7  
12:00 236.9 F 10.3  
14:00 236.2 F 8.8  
16:00 250.2 F 15.9  
18:00 262.6 F 23.0  
20:00 272.3 F 29.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.0 F 28.9  
1:00 272.5 F 29.6  
2:00 271.5 F 28.9  
3:00 271.6 F 29.4  
4:00 272.6 F 29.6  
  
Min 271.0 F 28.8  
Max 272.7 F 29.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 205.7 F 0.0  
5:00 214.4 F 0.0  
10:00 232.8 F 0.0  
15:00 234.6 F 0.0  
20:00 223.0 F 0.0  
25:00 220.6 F 0.0  
30:00 219.4 F 0.0  
Drying Complete

FILLING: 1:33  
HEATING: 20:14  
STERILIZING: 4:00  
VENTING: 3:09  
DRYING: 30:00  
TOTAL CYCLE: 00:58:56  
CYCLE COMPLETE

12/21/2022 12:17 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_