
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 003987
11:01 11/07/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:15
PRESSURIZING 2:22
STERILIZING 3:12
136.0 C 220kPa 3:12
136.1 C 220kPa 3:42
136.4 C 223kPa 4:12
136.3 C 222kPa 4:42
136.4 C 223kPa 5:12
136.4 C 222kPa 5:42
136.4 C 224kPa 6:12
136.1 C 225kPa 6:42
136.1 C 225kPa 6:42
VENTING 6:43
AIR DRYING 7:28
CYCLE COMPLETE 8:25

Digital Signature #
EE2466D1C971DABE
