

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2907

Name: RITTER2  
Sterilizer ID: V2313228  
09/20/2023 08:50 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 68.8 F 0.0  
2:00 75.6 F 0.0  
4:00 85.7 F 0.0  
6:00 102.3 F 0.0  
8:00 139.5 F 0.6  
10:00 166.8 F 1.7  
12:00 187.8 F 4.6  
14:00 219.1 F 8.4  
16:00 229.0 F 12.3  
18:00 226.3 F 6.2  
20:00 236.6 F 10.7  
22:00 233.5 F 7.8  
24:00 244.7 F 13.2  
26:00 242.8 F 11.9  
28:00 255.2 F 18.3  
30:00 266.0 F 24.9  
32:00 272.6 F 29.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.9 F 29.1  
1:00 272.4 F 29.0  
2:00 273.0 F 29.4  
3:00 273.0 F 29.7  
4:00 272.1 F 29.6  
  
Min 271.8 F 29.0  
Max 273.1 F 29.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 182.0 F 0.0  
5:00 228.2 F 0.0  
10:00 231.9 F 0.0  
15:00 199.9 F 0.0  
20:00 183.1 F 0.0  
25:00 187.0 F 0.0  
30:00 183.0 F 0.0  
  
Drying Complete

FILLING: 0:48  
HEATING: 32:22  
STERILIZING: 4:00  
VENTING: 3:18  
DRYING: 30:00  
TOTAL CYCLE: 01:10:28  
CYCLE COMPLETE

09/20/2023 10:00 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_