
Midmark M11 - v1.0.5
Total Cycles: 1364

Name: RITTER2
Sterilizer ID: V2313228
08/03/2022 01:29 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 183.7 F 0.0
2:00 174.1 F 0.3
4:00 191.2 F 3.1
6:00 207.8 F 7.7
8:00 224.2 F 12.6
10:00 230.7 F 8.9
12:00 238.9 F 10.6
14:00 244.1 F 12.9
16:00 243.5 F 12.2
18:00 256.1 F 18.9
20:00 267.4 F 25.9

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.8
1:00 272.2 F 29.2
2:00 271.3 F 28.9
3:00 272.2 F 29.3
4:00 271.4 F 28.6

Min 271.1 F 28.6
Max 272.4 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 197.4 F 0.0
5:00 231.6 F 0.0
10:00 208.2 F 0.0
15:00 217.2 F 0.0
20:00 235.3 F 0.0
25:00 222.2 F 0.0
30:00 226.7 F 0.0
Drying Complete

FILLING: 1:04
HEATING: 21:24
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 00:59:38
CYCLE COMPLETE

08/03/2022 02:28 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____