
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 008310
11:44 12/02/2025
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:34
PRESSURIZING 2:36
STERILIZING 3:41
136.4 C 222kPa 3:41
136.3 C 223kPa 4:11
136.4 C 223kPa 4:41
136.4 C 221kPa 5:11
136.4 C 223kPa 5:41
136.2 C 220kPa 6:11
136.3 C 223kPa 6:41
136.4 C 220kPa 7:11
136.4 C 220kPa 7:11
VENTING 7:12
AIR DRYING 8:05
CYCLE COMPLETE 38:05

Digital Signature #
DE46C3FE66760160
