
Brampton, ON
Canada

STATIM 5000 S501R713
SN 511019D00004
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001584
14:22 11/15/2021
UNWRAPPED
134 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:55
PRESSURIZING 6:44
STERILIZING 8:07
136.0 C 220kPa 8:07
136.0 C 219kPa 8:37
136.1 C 222kPa 9:07
135.8 C 218kPa 9:37
136.1 C 220kPa 10:07
136.4 C 222kPa 10:37
136.4 C 220kPa 11:07
136.4 C 223kPa 11:37
VENTING 11:37
AIR DRYING 12:44
CYCLE COMPLETE 13:51

Digital Signature #
EFAD08DD6DCAF766
