

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2592

Name: RITTER2  
Sterilizer ID: V2313228  
06/30/2023 07:45 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 72.5 F 0.0  
2:00 74.1 F 0.0  
4:00 96.7 F 0.0  
6:00 119.3 F 0.0  
8:00 149.4 F 0.0  
10:00 176.9 F 0.7  
12:00 191.9 F 4.2  
14:00 205.1 F 7.5  
16:00 217.4 F 10.9  
18:00 224.0 F 6.9  
20:00 232.2 F 9.4  
22:00 240.3 F 11.6  
24:00 241.9 F 11.8  
26:00 239.7 F 10.3  
28:00 252.2 F 16.5  
30:00 262.8 F 22.7  
32:00 272.6 F 29.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.5  
1:00 272.7 F 29.2  
2:00 271.6 F 28.3  
3:00 272.7 F 29.3  
4:00 272.1 F 28.7  
  
Min 271.5 F 28.3  
Max 272.8 F 29.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 175.7 F 0.0  
5:00 194.2 F 0.0  
10:00 196.4 F 0.0  
15:00 198.6 F 0.0  
20:00 196.6 F 0.0  
25:00 208.1 F 0.0  
30:00 208.7 F 0.0  
  
Drying Complete

FILLING: 0:52  
HEATING: 32:27  
STERILIZING: 4:00  
VENTING: 3:08  
DRYING: 30:00  
TOTAL CYCLE: 01:10:27  
CYCLE COMPLETE

06/30/2023 08:55 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_