

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4536

Name: RITTER 1  
Sterilizer ID: V2178198  
05/08/2024 04:05 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 109.1 F 0.0  
2:00 109.8 F 0.8  
4:00 121.2 F 2.7  
6:00 145.2 F 5.6  
8:00 167.4 F 8.9  
10:00 192.0 F 12.6  
12:00 206.6 F 5.5  
14:00 217.2 F 9.5  
16:00 227.1 F 11.5  
18:00 231.6 F 10.0  
20:00 236.7 F 9.0  
22:00 245.2 F 13.1  
24:00 255.6 F 19.0  
26:00 266.3 F 25.3  
28:00 272.2 F 29.1

STERILIZING  
mm:ss Degrees PSI  
0:00 272.1 F 29.1  
1:00 271.5 F 29.1  
2:00 271.9 F 29.4  
3:00 271.4 F 29.4  
4:00 271.4 F 29.5

Min 270.9 F 29.0  
Max 273.5 F 29.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 179.0 F 0.0  
5:00 236.7 F 0.0  
10:00 245.2 F 0.0  
15:00 220.6 F 0.0  
20:00 202.2 F 0.0  
25:00 188.8 F 0.0  
30:00 188.1 F 0.0  
Drying Complete

FILLING: 0:54  
HEATING: 28:01  
STERILIZING: 4:00  
VENTING: 3:44  
DRYING: 30:00  
TOTAL CYCLE: 01:06:39  
CYCLE COMPLETE

05/08/2024 05:11 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_